
Service information SI_2015_104_0073 T3

REPORTED SERVICE INFORMATION**0073 T3**

| Nr | Name | Description | Date |
|----|----------------------------|------------------------------------|-------------------|
| 01 | SI_2014_101_0073 T3 | Incorrect connection in 0055-table | 2014-11-03 |
| 02 | SI_2014_102_0073 T3 | Calibration of Autopositions | 2014-12-12 |
| 03 | SI_2014_103_0073 T3 | Grid detection input | 2014-12-18 |
| 04 | SI_2015_104_0073 T3 | New Software release | 2015-05-20 |
| | | | |
| | | | |

Service information SI_2015_104_0073 T3

| | | |
|---|--------------------------------------|--|
| System type: Precision systems T3 | Serial numbers: 2001-2047 | Date issued: 2015-05-20 |
| ServiceInfo number: SI_2015_104_0073 T3 | Issued by: Erik Säll | Approved by: Jonas Petersson |
| Time for action: WHEN CONVENIENT | Report back to Arcoma: N:a | Ref nr: (ÄBO, Complaint) N:a |
| Subject: New Software release | | |

1 INFORMATION

Arcoma have released a new software version for the 0072-T3 system. It is available for download at <http://arcoma-imix.com/wp-login.php>

The new software will be implemented in production from serial number 2048.

2 NEW FUNCTIONALITY

- Support for remote control handling.
- Collimator light will switch on when the auto position is reached.

3 ENHANCEMENTS AND BUG FIXES

- Tube display date format can now easily be selected by the user (YYYY-MM-DD, MM-DD-YYYY etc.).
- Patient age shown in tube display now corresponds fully with the age shown in the Canon Imaging System GUI.
- Accession number can be shown if needed.
- PostmAs value is removed from the display (still available on the Canon-CPI display).
- Table Stitching: Images included in the stitched image were earlier rotated for some system configurations while the stitched image was correct. Now the included images will be correct rotated for all configurations together with the stitched image. Please note that APR:s might need to be reviewed.

For more details of the update please see SW Release note.

4 COMPATIBILITY

Software validated for CPI CXDI NE-ETP v3.1.0.0.

For more details of the update please see SW Release note.



QUESTIONS

If you have any questions regarding this SI, please contact Arcoma service department on service@arcoma.se or +46 470 706970.

ATTENTION

Service information

The service information is an important way for ARCOMA to reach out to our dealers and their technicians. The information is a part of our CAPA-process (Corrective And Preventive Action) and this part is to correct problems or inform you about essential parts of this process, and prevent an incident or miss use to occur.

The CE-mark to MDD Class II products is fulfilled through MDD ANNEX 2 93/42/EEC including our Quality system, the CAPA-process is a part of our Quality system.

Please take this information serious and if there is any information that is not clear do not hesitate to contact us.

Arcoma service department sends out the service information and we communicate with our E-mail. Please confirm that you have received the e-mail and understand the content.

We use an e-mail group for all our dealers; so if you change e-mail or want to have it sent to another address, please inform us.

Best regards
ARCOMA AB

Service information guide

The information starts with a list of content of all sent Service information from Arcoma since the rutin was official. If any one is missed in your folder please let us know. The present SI is marked bold.

| | | |
|------------------------------|---|---|
| Stand type | : | Product that the service information concerns, name and/ or product id ex. 0070 ArcoCeil |
| Serial numbers | : | The product serial number, -s that is concerned. This can be an interval ex. 501 – 620. If no serial number mentioned it means all. |
| Date issued | : | The date the service information was sent out. |
| Service info number | : | The number of the service information. SI_YYYY-## |
| Issued by | : | The person creating and sending the information |
| Approved by | : | The person that has approved the information |
| Time for action | : | This information states the action that has to be carried out. |
| • <i>Urgent action</i> | : | Immediate action or closing the lab |
| • <i>As soon as possible</i> | : | Book the room at the hospital as soon as it is possible. If it is information to the hospital please send it at once. |
| • <i>When convenient</i> | : | At next service booked |
| • <i>Only if problems</i> | : | If a problem as described appear take action. |
| Report back | : | Report back with serial number when action has been performed |
| Subject | : | A header of the Service information |
| Ref nr: | : | Reference to internal or external reference number. Ex 13467, ÅBO_50430_01 |